



CR imaging Plate System

Certificate N°: **BAM/ZBF/006/16**
2nd Revised version



Bundesanstalt für
Materialforschung
und -prüfung

Hereby it is certified by the BAM Certification Body that the

Industrial CR Imaging Plate System

with the designation

SCANX DISCOVER HR using

Imaging Plates UH-IP

of the certificate holder

DÜRR NDT GmbH & Co. KG
74321 Bietigheim-Bissingen
Germany

12200 Berlin, Germany
T: +49 30 8104-0
F: +49 30 8104-7 2222

CERTIFICATE

meets the requirements of the highest system class IP 1 / 30 according to EN 14784-1:2005 and ISO 16371-1:2011 for industrial computed radiography with imaging plates for non-destructive testing, if the exposure dose is at minimum 15,7 mGy. This dose corresponds to a CEN speed and ISO speed of 64 at a scanner pixel size of 20 µm. CR system performance level II according to ASTM E 2446-16 is fulfilled at an ISO speed of 320. The maximum basic spatial resolution is 30 µm. Other system classes (CEN/ISO: IP 2 / 30 to IP 6 / 30) can be reached with lower exposure dose values (see test report No. 8.3/7830 of 2015-09-24). The spider net graph with the summary of the CR system characterization according to ASTM E 2446-16 is presented on the back-side of this certificate. Procedure N° BZS-GS/040/19 forms the basis of this 1st revision version.

The certification is performed on the basis of certification contract N° **BAM-ZBA-0001-2006-Dürr, 2nd revised version**, according to standard ISO/IEC 17065:2012 and comprises a design type test (BAM Certification System I).

The products certified by BAM may be labeled with the BAM certification mark „BAM Baumustergeprüft“ and/or “BAM Design-type tested” together with the certificate number.

The certificate is valid until 18 October 2023.

for Bundesanstalt für Materialforschung und -prüfung (BAM)
Unter den Eichen 87,12205 Berlin, **2020-01-20**

Dr. R. Schmidt
BAM Certification Body



Dipl.-Ing. B. Redmer
BAM Assessor

Distribution list: 1st Certificate holder

2nd BAM Certification Body

The BAM Certification Body has been accredited according to standard ISO/IEC 17065:2012 by the DAkkS (Deutsche Akkreditierungsstelle GmbH). The accreditation is valid for the scope given in certificate D-ZE-11075-21-00.


This certificate may only be published in full wording and without any additions. A revocable written consent shall be obtained from BAM beforehand for any amended reproduction or the publication of any excerpts. The German version is legally binding, except an English version is issued exclusively. Place of jurisdiction is Berlin.



Qualification of ScanX Discover HR with UH-IP imaging plates according to ASTM E 2446-16:



System parameters:
 ScanX Discover HR reader unit
 UH-IP imaging plates
 Scan resolution 20 μm
 BAM certified scanning mode



Annex to
 Certificate BAM/ZBF/006/16
 of 2020-01-20

Distribution list: 1st Certificate holder

2nd BAM Certification Body

The BAM Certification Body has been accredited according to standard ISO/IEC 17065:2012 by the DAkkS (Deutsche Akkreditierungsstelle GmbH). The accreditation is valid for the scope given in certificate D-ZE-11075-21-00.

This certificate may only be published in full wording and without any additions. A revocable written consent shall be obtained from BAM beforehand for any amended reproduction or the publication of any excerpts. The German version is legally binding, except an English version is issued exclusively. Place of jurisdiction is Berlin.